



**NATIONAL INSTITUTE OF PHARMACEUTICAL EDUCATION & RESEARCH**  
[Department of Pharmaceuticals, Ministry of Chemicals & Fertilizers, GoI]  
**HYDERABAD**

\*\*\*

**Amendment To Tender**

No: PUR/NIPERHYD/NC/30/19

December 6, 2019

To

Dear Sirs,

Sub: Extension of Tender for supply of **"HPLC (HIGH PERFORMANCE LIQUID CHROMATOGRAPHY) SYSTEM WITH PDA DETECTOR"**- reg.

Ref : Our Enquiry No: No: PUR/NIPERHYD/NC/30/19, Dt:13/11/19

We are incorporating the following changes in the above tender as per the following.

Sl. No.	As per the tender	Amended to read as
1.	Last date for submission of bids 11.12.2019 2.30 pm	Last date for submission of bids 01.1.2020 2.30 pm
2.	Opening of bids 11.12.2019 at 3.00pm	Opening of bids 1.1.2020 at 3.00pm

Note:

- 1.Please submit the bids as per the pre bid minutes uploaded in the website and also the tender specifications .
- 2.All other terms and conditions of the tender remain same.
3. All the bids are to be submitted in two bid system i.e. technical and commercial bids separately in different sealed covers. The EMD amount of Rs.1,59,000/- in the form of DD or bankers Cheque in favour of "NIPER-H Resources" should be kept in the technical bid cover to avoid its disqualification.

Thanking You,

Yours faithfully,

*yvm*  
9/12/19  
Dr. Y.V.Madhavi  
Chairman SPC-II



NATIONAL INSTITUTE OF PHARMACEUTICAL EDUCATION & RESEARCH  
[Department of Pharmaceuticals, Ministry of Chemicals & Fertilizers, GoI]  
HYDERABAD

\*\*\*

Tender No. PUR/NIPERHYD/NC/30/19 – HPLC SYSTEM WITH PDA DETECTOR

Subsequent to the pre bid conference held in the Institute in connection with procurement of HPLC System with PDA Detector against our tender No.PUR/NIPERHYD/NC/30/19, the Institute would like to add following information to the existing tender document.

1. The Institute would like to have additional 2 years comprehensive warranty coverage as an optional item. Suppliers are expected to quote separately for 2<sup>nd</sup> year and 3<sup>rd</sup> year cost implications towards additional warranty obligations. In case Institute desires to have additional warranty beyond 1 year the payment towards 2<sup>nd</sup> and 3<sup>rd</sup> year will be paid only after successful completion of warranty obligations each additional year separately.
2. As the Institute do not have any clearing agent, the Institute expects the suppliers to handle clearance of items from customs wherever applicable. However the Institute will provide required documents like duty exemption certificate, IEC Code and any other declaration to be given by Institute. The prices should be quoted in FOR upto NIPER Hyderabad Stores.
3. The list of following items quoted separately which may be added to the cost of main unit for the purpose of comparison. Depending on the budget availability Institute may select the required items from the following and only those selected items will be added for comparison purpose.

Following specifications will be added to the specifications of HPLC.

1. Auto sampler- Sample cooler /thermostat should have capability to cool to heat (around 5°C up to 40°C).
2. Comprehensive warranty: 3 years from date of installation (including computer and software's supplied)
3. Columns Optional items:
  - a. USP L1 (C18, 5 µm, 4.6 mm X 150 mm)-2 nos
  - b. USP L7 (C8, 5 µm, 4.6 mm X 150 mm)-1 nos
  - c. USP L11 (Phenyl, 5 µm, 4.6 mm X 150 mm)-2 nos
  - d. USP L3 (HILIC, 5 µm, 4.6 mm X 150 mm)-1 nos
  - e. USP L8 (Amino, 5 µm, 4.6 mm X 150 mm)-1 nos
  - f. USP L3 (Silica, 5 µm, 4.6 mm X 150 mm)-1 nos

The above additional information will be frame part of the tender document.

*your*  
*9/12/19*  
Dr. Y.V.Madhavi  
(Chairman)